

MAR 08 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>JEFFREY ELBIE</u>	ID# <u>6623</u> ✓ Date <u>2-24-11</u> ✓
A Agency <u>Anchorage Police Dept</u>	Phone # <u>786-8900</u>
Instrument Location <u>Anchorage Jail</u>	
B Alco S/N <u>78259</u> ✓ Target Value <u>.087</u> ✓ High Pressure <u>700</u>	
Alco Test Values <u>.091</u> ✓ 1st Alco <u>.091</u> ✓ 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 18th day of MAR, 2011 2011

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130579

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

FEBRUARY 24, 2011

OPERATOR'S NAME:

ELBIE/JEFFREY/L

OPERATOR'S NUMBER: 6623

SUBJECT'S LAST NAME:

TEST

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0000

DEPT/AGENCY: ANC1

CASE/REPORT: 11-00

TEST TYPE: V

ALCO TARGET VALUE: .087

ALCO S/N: 78259

--- BREATH ANALYSIS ---

.087 ADJUSTED FOR 30.97 in

ALCO TARGET .090

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 30.98 in .091

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 30.98 in .091

BLANK TEST .000

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130579

FEBRUARY 24, 2011

TIME 04:44

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 12/11/08

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 30.98 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJKLMN
OPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~Δ